# VENTILATOR CIRCUIT FOR O-TWO E700, E600 AND E500 ELECTRONIC TRANSPORT VENTILATORS

o_two controlled™ ventilation

REF 01CV8030-CS

⚠ CAUTION ⚠

1. Federal law restricts this device to sale by or on the order of a physician.
2. This device is only to be used with o-two e700, e600 and e500 Electronic Transport Ventilators.
3. Patient should be constantly monitored whenever this device is in use.
4. Store in clean dry conditions, away from heat and light.
5. The use of a filter with the disposable breathing circuit may increase inspiratory resistance
6. This circuit is designed for a single patient use only and cannot be disassembled. Do not attempt to clean, disinfect or re-use. Reusing this patient circuit may pose a risk of cross contamination while cleaning or disinfecting it may result in device malfunction.
7. Keep package sealed until use.
8. Dispose of patient circuit in accordance with local protocols.
9. Spare "o-two" patient circuits are recommended at all times.

DIRECTION FOR USE

1. Inspect circuit for clear air way passage prior to use.
2. Make sure connections are compatible and fit firmly.
3. Connect patient circuit as instructed in ventilators' user manual.
4. Perform leak test as outlined in ventilators instruction manual:
   a. Connect oxygen supply to input connector, test lung to the patient end of the circuit.
   b. Turn on ventilator and select child default setting.
   c. Press and hold manual button and observe the pressure wave form on the screen If pressure drops immediately, inspect patient circuit connections and ensure all connectors are attached. Replace circuit if necessary
5. Connect the patient circuit to the patient after the ventilator is turned on.

SPECIFICATIONS

Overall length	1.83 meter (6 feet)	Leakage rate	100 ml/min
Rated flow	80 L/min p<0.2 Kpa	Compliance	9.25 ml/Kpa/m

⊘2 LATEX NON STERILE -20°C / 50°C LOT ⧖

CE 1639